



Government of West Bengal
Office of the Principal
College of Medicine & Sagore Dutta Hospital
578, B.T.Road, Kamarhati, Kolkata-700058

Memo No: CMSDH/ Pr./2046/21

Date: 26-10-2021

Corrigendum Tender Notice

NIT No: CMSDH-14/2021-22

Tender ID: 2021_HFW_347019_1

Tender Name: Supply, Installation and Commissioning of Mid End Ultrasound Machine with Colour Doppler

Ref: Pre bid meeting dated 26.10.2021

Following points in technical specification are hereby modified:-

SI No	SI No in Specification	Existing	Modification
1	9	Frequency processing facility for the transducers should be 1 or 2 to 12 MHz. This must be available without the need for frequency switching.	Frequency processing facility for the transducers should be 1 or 2 to 12 MHz. This must be available with or without the need for frequency switching.
2	16	Acquisition frame rate of at least 1000 frames/second. Acquisition frame rate should be clearly mentioned in the technical quote.	Acquisition frame rate of at least 800 frames/second or more. Acquisition frame rate should be clearly mentioned in the technical quote.
3	20	Archive – facility to transfer images to an integrated CD writer, without any interfacing. Specify if integrated CD writer is available in your technical quote and USB drive facility.	Archive – facility to transfer images to an integrated CD writer, without any interfacing. Specify if integrated CD writer is available in your technical quote and/or USB drive facility.
4	21	Print-Should have direct connectivity to Inkjet printer for printing images & report.	Print-Should have direct or indirect connectivity to Inkjet printer for printing images & report.
5		(Addition in technical Specification)	The system should have 14" gesture control touch screen facility
6		(Addition in technical Specification)	System should have a battery backup at least 2 hours or more

Dhor 26/10/2021

Principal
College of Medicine & Sagore Dutta Hospital
Kamarhati, Kolkata-700058